



Automatikprodukter

CALIBRATION PROTOCOL

Order No 6061 Date 2015-10-05
 Customer Protektor i Sundsvall AB
 Article Number SNH3 1000

Reference dates

Calibration gas NH3 Alternate calibration
 Gas concentration 500 ppm
 % LEL
 % vol
 Measure range 30-1000 ppm
 % LEL
 % vol
 Temperature 21° C Humidity 45 %
 Pressure 1000 mbar Warm-up time sensor 192 h
 Given calibration value 120 mV Actual calibration value 120 mV
 Difference given/actual value %
 Gas flow 150 l/min

Option relays

Threshold Relay 1 (ppm, % LEL/ % vol) Hysterese
 Threshold Relay 2 Manufacturing number 13976
 Threshold Relay 3 13979
 Threshold Relay 4

..... Ann Margreth Johansson

Ewert Johansson/Ann-Margreth Johansson

Address:
 Automatikprodukter i Askim AB
 Gåsossen 43
 SE-436 39 ASKIM
 Sweden

Phone:
 Nat 031-28 72 02
 Int +46 31 28 72 02
 Web: www.automatikprodukter.se
 Styrelsens säte: Göteborg

VAT.No: SE556301392801
 Innehar F-skattsedel
 EORI No. SE5563013928



Automatikprodukter

CALIBRATION PROTOCOL

Order No 6061 Date 2015-10-05
 Customer Protektor i Sundsvall AB
 Article Number SNH3 1000

Reference dates

Calibration gas NH3 Alternate calibration
 Gas concentration 500 ppm
 % LEL
 % vol
 Measure range 30-1000 ppm
 % LEL
 % vol
 Temperature 21° C Humidity 45 %
 Pressure 1000 mbar Warm-up time sensor 192 h
 Given calibration value 120 mV Actual calibration value 120 mV
 Difference given/actual value %
 Gas flow 150 l/min

Option relays

Threshold Relay 1 (ppm, % LEL/ % vol) Hysterese
 Threshold Relay 2 Manufacturing number 13976
 Threshold Relay 3 13979
 Threshold Relay 4

..... Ann Margreth Johansson
 Ewert Johansson/Ann-Margreth Johansson

Address:
 Automatikprodukter i Askim AB
 Gåsossen 43
 SE-436 39 ASKIM
 Sweden

Phone:
 Nat 031-28 72 02
 Int +46 31 28 72 02
 Web: www.automatikprodukter.se
 Styrelsens säte: Göteborg

VAT.No: SE556301392801
 Innehar F-skattsedel
 EORI No. SE5563013928



Automatikprodukter

CALIBRATION PROTOCOL

Order No 6061 Date 2015-10-05
 Customer Paolator i Sundsvall AB
 Article Number SNH3 1000

Reference dates

Calibration gas NH3 Alternate calibration
 Gas concentration 500 ppm
 % LEL
 % vol
 Measure range 30-1000 ppm
 % LEL
 % vol
 Temperature 21° C Humidity 45 %
 Pressure 1000 mbar Warm-up time sensor 192 h
 Given calibration value 120 mV Actual calibration value 120 mV
 Difference given/actual value %
 Gas flow 150 l/min

Option relays

Threshold Relay 1 (ppm, % LEL/ % vol) Hysterese
 Threshold Relay 2 Manufacturing number 13976
 Threshold Relay 3 13979
 Threshold Relay 4

..... Ann Margreth Johansson
 Ewert Johansson/Ann-Margreth Johansson

Address:
 Automatikprodukter i Askim AB
 Gåsmossen 43
 SE-436 39 ASKIM
 Sweden

Phone:
 Nat 031-28 72 02
 Int +46 31 28 72 02
 Web: www.automatikprodukter.se
 Styrelsens säte: Göteborg

VAT.No: SE556301392801
 Innehar F-skattsedel
 EORI No. SE5563013928